

# Certification Record



Listing#: E114778  
Report #: 101011  
Original Certification: January 31, 2019  
Revised Certification: N/A

This Certification is issued to:

SynTouch  
3720 Clifton Pl.  
Montrose, CA 91020  
USA

For the product(s):  
Haptic Measurement Instrument,  
Model Toccare

Has been certified to the following standard(s):  
UL61010-1, Third Edition: Safety of Electrical Equipment For Measurement, Control, and Laboratory Use, Rev. April 29, 2016.  
UL61010-2-81, Second Edition: Particular Requirements for Automatic and Semi-Automatic Laboratory Equipment for Analysis and Other Purposes, Rev. August 14, 2015.

A handwritten signature in black ink, appearing to read "Nehemya Cohen".

Nehemya Cohen  
Project Engineer,  
MET Safety Laboratory

*All changes proposed in the previously identified product that affects the above information must be submitted to MET for evaluation prior to implementation to assure continued MET Certification status.*

*The covered product(s) shall be subject to follow-up inspections to ensure that the Certified product(s) are identical to the product sample evaluated by MET Laboratories, Inc. and that all manufacturer's responsibilities are being fulfilled as specified in the Manufacturer's Responsibility section of the Certification report. The applicant named above has been authorized by MET Laboratories, Inc. to represent the product(s) listed in this record as "MET Certified" and to mark this/these product(s) according to the terms and conditions of the MET Applicant Contract, MET Listing Reports, and the applicable marking agreements. Only the product(s) bearing the MET Mark and under a follow-up service are considered to be included in the MET Certification program. This certification has been granted under a System 3 program as defined in ISO/IEC 17067.*



*MET Laboratories, Inc. is accredited by OSHA and the Standards Council of Canada.  
The Nation's First Nationally Recognized Testing Laboratory*

**NRTL**